

Notice of Allowability

Application No.

10/801,830

Examiner

John Kim

Applicant(s)

CHEVALLET ET AL.

Art Unit

1723

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to 7/20/07.
2. ☒ The allowed claim(s) is/are 1-38.
3. ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☒ All b) ☐ Some* c) ☐ None of the:
 1. ☒ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

* Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.
THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
 5. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
 - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

1. ☐ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
3. ☐ Information Disclosure Statements (PTO/SB/08),
Paper No./Mail Date _____
4. ☐ Examiner's Comment Regarding Requirement for Deposit
of Biological Material
5. ☐ Notice of Informal Patent Application
6. ☐ Interview Summary (PTO-413),
Paper No./Mail Date _____
7. ☒ Examiner's Amendment/Comment
8. ☐ Examiner's Statement of Reasons for Allowance
9. ☐ Other _____

Art Unit: 1723

1. An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with 8/8/07 on Aaron Parker.

The application has been amended as follows:

-Abstract has been replaced with new abstract on following separate page.

-Claim 38, line 11: Between "first" and "outlet", -- fluid -- has been inserted.

line 18: Between "second" and "outlet", -- fluid -- has been inserted.

line 19: Between "first" and "inlet", -- fluid -- has been inserted.

line 28: Between "first" and "inlet", -- fluid -- has been inserted.

line 32: Between "the" and "output", -- blood -- has been inserted.

2. Any inquiry concerning this communication or earlier communications from the examiner should be directed to /John Kim/ whose telephone number is 571-272-1142. The examiner can normally be reached on Monday-Friday 7 a.m. - 3:30 p.m..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, David Sample can be reached on 571-272-1376. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 1723

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

**/John Kim/
Primary Examiner
Art Unit 1723**

JK
8/8/07

--

Abstract of the Disclosure

A device and method for an extracorporeal blood treatment is described. The device comprising at least one exchanger equipped with at least one first inlet for the blood to be treated, a first fluid outlet and a second fluid outlet, an input line for blood to be treated connected to the first inlet of the exchanger, a blood output line (or venous line) connected to the first outlet of the exchanger, at least one treatment unit comprising at least one first fluid inlet and at least one first fluid outlet, the second outlet of the exchanger being in fluid communication with the first inlet of the treatment unit, where the first outlet from the treatment unit is in fluid communication with the input line and the second outlet of the treatment unit is in fluid communication with a first waste liquid discharge line.--.